



California Medical Device Recall Information



Recall Name

Bestmed Recalls Digital Temple Thermometers Due to Inaccurate Temperature

Recall Date	Product Description	Recalling Firm	Recall Reason
11/12/15	Digital Temple Thermometer “DTT” Model No. KD-2201	Distributed by: Bestmed, LLC. Golden, CO Manufactured by: K-Jump Health Co., Ltd.	<i>Some thermometers are inaccurate and display temperatures lower than actual body temperatures.</i>
Recall Class	Product Identification	Distribution	Affected Dates
II	<ul style="list-style-type: none">• S/N: 3612 through 3715• Distributed under the following brand names: Bestmed Good Neighbor Kroger Medline Meijer Premier Value Safeway Life Brand Target Top Care Best Choice (Value Merchandise) Western Family	Nationwide	Sold between: October 2012 to November 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm474774.htm>